

Certificate of Analysis

Product Name	Cell Counting Kit-8 (CCK-8)		
Cat.No.	C6005M	Storage Conditions	4℃&Protect from light
Lot.No.	211104Z01-05	Expiration Date	3-11-2024
Molecular Formula	N/A	Molecular Weight	N/A
CAS Number	N/A	Excitation Emission	N/A
Quality Control			
Items	Specification	Result	
Packaging appearance	Intact and undamaged	Pass	
Appearance	Red solution	Pass	
Weight/volume	5-5.25mL	Pass	
Appendix			
This Product has been tested and the components have been found to conform to UE quality standards.			
Conclusion	Up to standard	Report Date	6-1-2022
QC	zhen wei 6-1-2022	QA	Qin hua.Yu 06-01-2022



Appendix

Standard

Cell	Six holes						Avg	Subtract blank
20000	2.0965	1.9669	1.9502	2.1585	1.7569	1.8747	1.967283	1.815483
10000	1.0905	1.0849	1.0924	0.9348	0.948	1.0177	1.02805	0.87625
5000	0.6301	0.6325	0.6437	0.5469	0.652	0.5227	0.60465	0.45285
2500	0.361	0.3556	0.3424	0.3986	0.4918	0.3232	0.378767	0.226967
1250	0.2473	0.242	0.2491	0.1945	0.2164	0.224	0.228883	0.077083
625	0.1979	0.1937	0.1987	0.2022	0.2043	0.1637	0.193417	0.041617
312.5	0.1748	0.1715	0.1736	0.1743	0.179	0.1731	0.174383	0.022583
0	0.152	0.152	0.1519	0.1512	0.1521	0.1516	0.1518	0

This lot

Cell	Six holes						Avg	Subtract blank
20000	1.9166	2.1442	2.2044	2.0596	1.9693	2.0266	2.05345	1.8994
10000	0.9688	0.967	1.0583	1.1586	1.1387	1.1494	1.073467	0.919417
5000	0.5401	0.5865	0.6517	0.485	0.5295	0.635	0.5713	0.41725
2500	0.3354	0.3461	0.3476	0.4144	0.4393	0.3723	0.37585	0.2218
1250	0.2484	0.2575	0.2601	0.2246	0.2468	0.2286	0.244333	0.090283
625	0.1926	0.1994	0.1818	0.2055	0.1757	0.1781	0.18885	0.0348
312.5	0.1711	0.1752	0.174	0.1678	0.181	0.1758	0.17415	0.0201
0	0.1548	0.1528	0.1574	0.1536	0.1527	0.153	0.15405	0

