

Certificate of Analysis

Product Name	Cell Counting Kit-8 (CCK-8)		
Cat.No.	C6005MKT	Storage Conditions	4℃&Protect from light
Lot.No.	211104Z01-07	Expiration Date	3-11-2024
Molecular Formula	N/A	Molecular Weight	N/A
CAS Number	N/A	Excitation Emission	N/A
Quality Control			
Items	Specification	Result	
Packaging appearance	Intact and undamaged	Pass	
Appearance	Red solution	Pass	
Weight/volume	5-5.25 mL	Pass	
Appendix			
This Product has been tested and the components have been found to conform to UE quality standards.			
Conclusion	Up to standard	Report Date	21-1-2022
QC	Zhen Wei 21-1-2022	QA	Bia Hua Yu 21-01-2022



Appendix

Standard

Cell	Six holes						Avg	Subtract blank
20000	2.9872	3.1204	3.0884	2.9326	3.0756	2.9083	3.01875	2.883033
10000	1.86	1.8035	1.8234	1.7648	1.8686	1.8278	1.824683	1.688967
5000	1.0329	1.0887	1.0577	0.9601	0.9649	0.9607	1.010833	0.875117
2500	0.5385	0.5465	0.5326	0.531	0.5623	0.542	0.54215	0.406433
1250	0.3401	0.3362	0.3451	0.348	0.3644	0.3033	0.339517	0.2038
625	0.2293	0.2304	0.2267	0.2398	0.2312	0.2334	0.2318	0.096083
312.5	0.1862	0.185	0.1888	0.1843	0.1793	0.1763	0.183317	0.0476
0	0.1398	0.1385	0.1396	0.1358	0.1354	0.1252	0.135717	0

This lot

Cell	Six holes						Avg	Subtract blank
20000	2.977	3.0079	3.004	3.0017	2.9195	2.9102	2.97005	2.845133
10000	1.8271	1.8384	1.8144	1.8733	1.7552	1.6716	1.796667	1.67175
5000	1.0342	1.0261	1.051	1.0139	0.9817	0.8831	0.998333	0.873417
2500	0.5575	0.5226	0.5182	0.5564	0.5564	0.48	0.53185	0.406933
1250	0.3359	0.3305	0.3423	0.2993	0.282	0.2892	0.3132	0.188283
625	0.2337	0.2277	0.2321	0.2238	0.2249	0.2213	0.22725	0.102333
312.5	0.1819	0.187	0.1887	0.1775	0.1688	0.1725	0.1794	0.054483
0	0.1402	0.1391	0.1371	0.1124	0.1112	0.1095	0.124917	0

