


Certificate of Analysis

Product Name	TBE Buffer, 5×		
Cat.No.	211115L01-02	Storage Conditions	Room temperature
Lot.No.	DZ8040Z1	Expiration Date	14-11-2022
Molecular Formula	N/A	Molecular Weight	N/A
CAS Number	N/A	Excitation Emission	N/A
Quality Control			
Items	Specification	Result	
Packaging appearance	Intact and undamaged	Pass	
Appearance	Colorless solution	Pass	
Weight/volume	1-1.02L	Pass	
pH (25℃)	8.3±0.2	Pass	
Appendix			
			
This Product has been tested and the components have been found to conform to US Everbright quality standards.			
Conclusion	Up to standard	Report Date	06-12-2021
QC	Zhen Wei 6.12.2021	QA	Qin hua. Yu 06-12-2021

