

Certificate of Analysis

Product Name	Qbtest™ X-Green II dsDNA Quantitation Kit Plus		
Cat.No.	Q2038LT	Storage Conditions	Store at 4°C away from light
Lot.No.	230206L01-01	Expiration Date	05-08-2024
Molecular Formula	N/A	Molecular Weight	N/A
CAS Number	N/A	Excitation Emission	Part A: 480/520 nm (combine dsDNA)
Quality Control			
Items	Specification	Result	
Packaging appearance	Intact and undamaged	Intact and undamaged	
Appearance	A: Orange liquid B: Colorless liquid	A: Orange liquid B: Colorless liquid	
	C: Colorless liquid D: Colorless liquid	C: Colorless liquid D: Colorless liquid	
Weight/volume	A:1.25-1.3125mL B:250-255mL	A:1.25mL B:252mL	
	C:5-5.25mL D:5-5.25mL	C:5.1mL D:5.1mL	
CV	≤10%	4.6%	
Appendix			
This Product has been tested and the components have been found to conform to UE quality standards.			
Conclusion	Up to standard	Report Date	13-02-2023
QC	Jiaxing Wu	QA	Qin Hua-Yu

