

Certificate of Analysis

Product Name	Qbtest®1× dsDNA HS Assay Kit		
Cat.No.	Q2061S	Storage Conditions	Store at 4°C away from light
Lot.No.	220726L01-02	Expiration Date	25-07-2023
Molecular Formula	N/A	Molecular Weight	N/A
CAS Number	N/A	Excitation Emission	Part A: 480/520 nm (combine dsDNA)
Quality Control			
Items	Specification		Result
Packaging appearance	Intact and undamaged		Pass
Appearance	A: Light red liquid B: Colorless liquid C: Colorless liquid		Pass
Weight/volume	A:50-52.5ml B:1-1.05ml C:1-1.05ml		Pass
CV	≤10%		Pass
Appendix			
This Product has been tested and the components have been found to conform to UE quality standards.			
Conclusion	Up to standard	Report Date	10-08-2022
QC	Long Zhang	QA	Qin Hua-Yu

