



Certificate of Analysis

UElandy Inc.

Room 302, Building 1, No.18& Chunyao Road,

Xiangcheng District, Suzhou,

Jiangsu Province P.R, 215143 China

www.uelandy.com

Refer to website for regional contact information.

Product Name:	UE Multisource Total RNA Miniprep Kit	Lot ID:	230523KD1
Catalog Number:	UE-MN-MS-RNA-50	Expiration Date:	11/22/2024

Quality Control

Spectrophotometric Analysis


- Acceptable Range for $A_{260}/A_{280}=2.0\pm0.15$.
- Comparison of the Test Sample to a previously tested Standard. Acceptable range for Test Sample is $\leq \pm 10\%$ of Standard, based upon spectrophotometric analysis. The "Standard" is a previously tested and validated sample.

Electrophoretic Analysis

- Agarose gel electrophoresis of the samples exhibits a tight genomic DNA band without significant visible degradation.
- Agarose gel electrophoresis of the samples exhibits no contaminating RNA.

Gel Image and Test Data

Item	Test Date						Sample
	1	2	3	4	5	6	
A_{260}	1.794	1.829	1.857	2.094	1.769	1.659	1,2,3 are duplicate Test Samples 4,5,6 are duplicate Standards
A_{280}	0.853	0.874	0.886	0.979	0.863	0.799	
A_{260}/A_{280}	2.102	2.092	2.095	2.140	2.049	2.075	
Yield (μg)	25.110	25.604	25.998	29.322	24.761	23.223	
Electrophoresis	√	√	√	√	√	√	PASS
No visible RNA	√	√	√	√	√	√	
Endonuclease digest	√	√	√	√	√	√	
[Test Sample/Standard] Yield Difference	-0.768%						PASS



UElandy Inc. warrants and represents that all inspections and testing for the article described above have been performed at UElandy Inc. QC laboratory located on-site at the manufacturing plant. UElandy Inc. manufacturing facility is located at Room 302, Building 1, No.18& Chunyao Road, Xiangcheng District, Suzhou, Jiangsu Province P.R, 215143 China

(QC Inspector) *tian qin zheng*

(Approver) *Qin Hua-Yu*