

Certificate of Analysis

UElany Inc.

Room 302, Building 1, No.18& Chunyao Road,

Xiangcheng District, Suzhou,

Jiangsu Province P.R, 215143 China

www.uelandy.com

Refer to website for regional contact information.

Product Name:

UE Plasmid Miniprep Kit

Lot ID:

220425KA1

Catalog Number:

UE-MN-P-250

Expiration Date:

04/24/2025

Quality Control

Spectrophotometric Analysis

- Acceptable Range for A260/A280=1.8±0.15.
- Comparison of the Test Sample to a previously tested Standard. Acceptable range for Test Sample is ≤±10% of Standard, based upon spectrophotometric analysis. The "Standard" is a previously tested and validated sample.

Restriction Analysis


• lμg of plasmid is digested to completion by EcoRI

Electrophoretic Analysis

- Agarose gel electrophoresis of the samples exhibits a defined plasmid band with no degradation.
- Agarose gel electrophoresis of the samples exhibits no contaminating RNA.
- Agarose gel electrophoresis of the samples exhibits no contaminating bacterial genomic DNA.

Gel Image and Test Data

Item	Test Date						Comm1a
	1	2	. 3	4	5	6	Sample
A260	0.320	0.360	0.368	0.364	0.323	0.329	
A280	0.164	0.199	0.202	0.201	0.169	0.169	1,2,3 are duplicate Test Samples 4,5,6 are duplicate Standards
A260/A280	1.944	1.811	1.821	1.808	1.911	1.946	
Yield (µg)	4.797	5.401	5.521	5.461	4.852	4.940	
Electrophoresis	1	√	√	√	√	√	
No visible RNA	1	1	1	1	1	1	PASS
Endonuclease digest	1	√	√	√	√	√	
[Test Sample/Standard]	3.06%						PASS
Yield Difference							



*1 and 2 respectively represent the test and standard enzyme digestion

UElany Inc. warrants and represents that all inspections and testing for the article described above have been performed at UElany Inc. QC laboratory located on-site at the manufacturing plant. UElany Inc. manufacturing facility is located at Room 302, Building 1, No.18& Chunyao Road, Xiangcheng District, Suzhou, Jiangsu Province P.R, 215143 China

(QC Inspector)

Min Ma
$$11-5-20>2$$

(Approver)

 $314444444413-5-20>2$