



# Certificate of Analysis

UElandy Inc.

Room 302, Building 1, No.18& Chunyao Road,

Xiangcheng District, Suzhou,

Jiangsu Province P.R, 215143 China

www.uelandy.com

Refer to website for regional contact information.

Product Name:	UE PCR Clean-up Kit	Lot ID:	221125KB1
Catalog Number:	UE-PCR-250	Expiration Date:	11/24/2025

## Quality Control

### Electrophoretic Analysis

- Comparison of the Test Sample to a previously tested Standard. Acceptable range for Test Sample is  $\leq \pm 10\%$  of Standard, based upon analysis. The “Standard” is a previously tested and validated sample.
- The recovery of DNA from agar gel is more than 70%.

### Recovery Analysis

- The recovery of DNA from agar gel is more than 70%. Loading 8  $\mu$ l UE-PCR product mixture to the 2% agar gel for electrophoresis, and also loading 5.6  $\mu$ l and 8  $\mu$ l 100bp Ladder for 70% and 100% as contrast to estimate the recovery of purified PCR product by visual examination.

### Gel Image and Test Data

Item	Test Date						Sample
	1	2	3	4	5	6	
Yield (%)	100	100	100	100	100	100	1,2,3 are Test Samples 4,5,6 are Standards
[Test Sample/Standard] Yield Difference	0%						PASS

**100%M    STANDARD    TEST    70%M**



UElandy Inc. warrants and represents that all inspections and testing for the article described above have been performed at UElandy Inc. QC laboratory located on-site at the manufacturing plant. UElandy Inc. manufacturing facility is located at Room 302, Building 1, No.18& Chunyao Road, Xiangcheng District, Suzhou, Jiangsu Province P.R, 215143 China

(QC Inspector) *tian qin zheng*

(Approver) *Qin Hua-Yu*