



Certificate of Analysis

| | | | |
|---|-------------------------------------|--|--|
| Product Name | FITC-Annexin V and PI Apoptosis Kit | | |
| Cat.No. | F6012LM | Storage Conditions | Keep away from light at 4°C |
| Lot.No. | 240422L01-03 | Expiration Date | 21-04-2027 |
| Molecular Formula | N/A | Molecular Weight | N/A |
| CAS Number | N/A | Excitation Emission | B: Ex/Em = 494/518 nm C: Ex/Em = 535/617 nm (with DNA) |
| Quality Control | | | |
| Items | Specification | | Result |
| Packaging appearance | Intact and undamaged | | Intact and undamaged |
| Appearance | A: Colorless liquid | | A: Colorless liquid |
| | B: Colorless liquid | | B: Colorless liquid |
| | C: Red liquid | | C: Red liquid |
| Weight/volume | A:50-52.5ml B:250-265μl | | A:50ml B:250μl |
| | C:500-530μl | | C:500μl |
| Appendix | | | |
| STANDARD | | TEST | |
|  | |  | |
| This Product has been tested and the components have been found to conform to UE quality standards. | | | |
| Conclusion | Up to standard | Report Date | 18-03-2025 |
| QC | Cen ming Jiang | QA | Qin Hua-Yu |

