

Certificate of Analysis

Product Name	PVDF, 0.45 μ m, 265mm*3750mm		
Cat.No.	PF0045	Storage Conditions	Sealed storage at room temperature to avoid moisture
Lot.No.	250728E01-01	Expiration Date	27-07-2030
Molecular Formula	N/A	Molecular Weight	N/A
CAS Number	N/A	Excitation Emission	N/A
Quality Control			
Items	Specification	Result	
Packaging appearance	Intact and undamaged	Intact and undamaged	
Appearance	The surface of the filter material should be clean, flat, intact, without mechanical damage, and without fiber shedding; The sheet should have no obvious burrs or brushed edges.	Pass	
Bubble-point pressure	The alcohol bubble-point pressure shall be within the range of 0.035-0.1 MPa.	0.066-0.073MPa	
Water flow rate	The water flow rate shall be no less than 12,000 LMH/bar	21069-30231LMH/Bar	
Thickness	0.12-0.2mm	0.135-0.176mm	
Appendix			
This Product has been tested and the components have been found to conform to UE quality standards.			
Conclusion	Up to standard	Report Date	11-09-2025
QC	<i>Cen ming Jiang</i>	QA	<i>Qin Hua-Yu</i>

